



**HAL**  
open science

## Minimum Reporting Standards Should be Expected for Preclinical Radiobiology Irradiators and Dosimetry in the Published Literature

Yannick Poirier, Larry Dewerd, François Trompier, Morgane dos Santos, Ke Sheng, Keith Kunugi, Merriline Satyamitra, Andrea Dicarolo, Andrea Dicarolo, Thomas Winters

### ► To cite this version:

Yannick Poirier, Larry Dewerd, François Trompier, Morgane dos Santos, Ke Sheng, et al.. Minimum Reporting Standards Should be Expected for Preclinical Radiobiology Irradiators and Dosimetry in the Published Literature. *Radiation Research*, 2023, 200 (3), pp.217-222. 10.1667/RADE-23-00119.1 . irsn-04207658

**HAL Id: irsn-04207658**

**<https://irsn.hal.science/irsn-04207658>**

Submitted on 14 Sep 2023

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Copyright

1 **Commentary: Minimum Reporting Standards Should be Expected for Preclinical Radiobiology**

2 **Irradiators and Dosimetry in the Published Literature**

3 Yannick Poirier<sup>1</sup>, Larry A. DeWerd<sup>2</sup>, François Trompier<sup>3</sup>, Morgane Dos Santos<sup>4</sup>, Ke Sheng<sup>5</sup>, Keith Kunugi<sup>2</sup>,  
4 Merriline M. Satyamitra<sup>6</sup>, Andrea L. DiCarlo<sup>6</sup>, Thomas A. Winters<sup>a6</sup>

5 <sup>1</sup>Department of Radiation Oncology, University of Maryland School of Medicine, Baltimore, Maryland,  
6 USA

7 <sup>2</sup>Medical Radiation Research Center, Department of Medical Physics, University of Wisconsin School of  
8 Medicine and Public Health, Madison, Wisconsin, USA

9 <sup>3</sup>Ionizing Radiation Dosimetry Laboratory (LDRI), Human Radiation Protection Unity, Institut de  
10 Radioprotection et de Sûreté Nucléaire (IRSN), Fontenay-aux-Rose, France

11 <sup>4</sup>Department of Radiobiology and Regenerative Medicine (SERAMED), Radiobiology of Accidental  
12 Exposure Laboratory (LRAcc), Institut de Radioprotection et de Sûreté Nucléaire (IRSN), Fontenay-  
13 aux-Roses, France

14 <sup>5</sup>Department of Radiation Oncology, University of California San Francisco, San Francisco, California,  
15 USA

16 <sup>6</sup>Radiation and Nuclear Countermeasures Program (RNCP), Division of Allergy, Immunology and  
17 Transplantation (DAIT), National Institute of Allergy and Infectious Diseases (NIAID), National  
18 Institutes of Health (NIH), Rockville, Maryland, USA

19 <sup>a</sup> Corresponding Author

20 5601 Fishers Lane, Room 7A65

21 Rockville, MD 20852

22 301-761-7918

23 twinters@mail.nih.gov

## 24 **Introduction**

25           Preclinical researchers have long acknowledged the critical need for accurate dosimetry in the  
26 conduct of radiation-exposure studies involving living systems. Although radiation dosimetry is required to  
27 be National Institute for Standards and Technology (NIST)-traceable for human irradiation protocols, there  
28 were previously only limited requirements stipulating accurate radiation exposures for preclinical radiation  
29 studies funded by the National Institutes of Health (NIH). In one of the earliest identified publications  
30 (1956) that addressed radiation dosimetry in murine studies, Hurst *et al.* (1) discussed their calculations  
31 surrounding estimating neutron responses, and “distribution of of absorbed energy versus LET”. In 2009,  
32 Zoetelief *et al.* (2) outlined European efforts to harmonize X-ray dosimetry studies in radiation biology,  
33 which resulted in the establishment of a protocol to improve dosimetry oversight and allow for better  
34 comparison of results across different institutions. It is, however, unclear if these early efforts substantially  
35 changed the approach of biologists to ensuring accurate and reproducible dosimetry in their radiation  
36 research.

37           The lack of consideration of dosimetry in *in vitro* and *in vivo* work was further explored in 2011  
38 during the “Radiation Dose is More than a Number” workshop convened by NIST, and co-sponsored by the  
39 National Institute of Allergy and Infectious Diseases (NIAID), and the National Cancer Institute (NCI). The  
40 report that followed (3) made nine recommendations that ranged from coordinating with radiation physicists  
41 during the experimental design phase and including more setup and dosimetry details in publications, to  
42 establishing working groups to develop protocols, and implementing formal dosimetry intercomparison  
43 programs. In an attempt to address the latter suggestion, and ensure rigor and reproducibility in their funded  
44 research portfolio consistent with NIH grants policy<sup>1</sup>, the NIAID Radiation and Nuclear Countermeasures  
45 Program (RNCP) released a request for proposals (NIAID-NIH-RFP-NIHAI201800020<sup>2</sup>) in 2019, seeking  
46 to make a single contract award for “RNCP-Wide Dosimetry Guidance & Monitoring of Sources and  
47 Irradiation Protocols”. This funding opportunity requested that offerors develop a consistent means of  
48 dosimetry comparison and reproducibility, as well as provide the administrative foundation necessary to

---

<sup>1</sup> <https://grants.nih.gov/policy/reproducibility/index.htm>

<sup>2</sup> <https://sam.gov/opp/360d070bdf5d437f83e88c13442fa23f/view>

49 facilitate and coordinate dosimetry activities in partnership with the NIAID. Respondents were asked to  
50 provide services, facilities, expertise, and capabilities to develop a centralized dosimetry harmonization  
51 effort that fits the resources and circumstances of projects across the funded RNCP portfolio. The contract  
52 award was made in 2020 to the University of Wisconsin (Principal Investigator Larry DeWerd), who  
53 established a harmonization protocol to encompass all the irradiators in use across the RNCP-funded  
54 portfolio of grants, contracts, and inter-agency agreements.

55 In 2019, the RNCP also became concerned with the specifics of establishing and reporting neutron  
56 dosimetry and convened a “Neutron Radiobiology and Dosimetry Workshop” with presentations from 17  
57 subject matter experts, planned in collaboration with the Department of Defense (DoD), Defense Threat  
58 Reduction Agency (DTRA), and the National Aeronautics and Space Administration (NASA). The meeting  
59 report that was published in 2021 (4) explored historical neutron exposure research and highlighted the need  
60 for harmonized reporting of exposure and experimental parameters. The authors concluded that efforts  
61 should be made to update computer coding for neutron dosimetry estimates, government policies needed to  
62 be re-visited in terms of expected radiation energy spectrums, and funding agencies must make key  
63 investments in consideration of neutron dosimetry, to improve science and ensure radiation emergency  
64 preparedness.

65 All of these early attempts to ensure the use of accurate dosimetry in research laboratories  
66 culminated in the current effort to take the next step and make strong recommendations that sufficient  
67 information be included in manuscripts or supplemental data files, such that reviewers and editors can be  
68 assured that the irradiations were carried out correctly, and so readers of the published work have the  
69 necessary information to replicate published experiments.

70

## 71 **Background and Rationale for Dosimetry Reporting**

72 Good science must be well described to be understood, compared, and replicated. For radiation  
73 biology, this requires the combination of both biology and radiation physics – two very different expertises,  
74 typically provided by different individuals with specialized training. While biological systems can be

75 difficult to control, particularly for *in vivo* models, radiation follows well-understood physical processes,  
76 which can be controlled and measured to a very high level of accuracy, precision, and reproducibility.  
77 Therefore, it is both feasible, and important, to require radiobiological studies to describe their irradiation  
78 protocol.

79 Unfortunately, experimental irradiation protocols are currently not well described in the majority of  
80 published radiobiology studies. This was already true in 2011, during the original NIST workshop (3), and  
81 has since been confirmed by multiple meta-studies showing no apparent improvements in reporting since  
82 the original symposium (5-9). In a recent analysis of 1758 published pre-clinical studies employing ionizing  
83 radiation, some parameters, such as radiation dose, are nearly always provided (>94%). Many important  
84 details, such as the equipment producing the radiation, the radiation type, beam energy (or quality), and dose  
85 rate, are usually, but not always, reported (>50-80%). Details pertaining to experimental design, such as the  
86 number and geometry of the radiation fields, the use of shaping collimators, or rotating tables to ensure  
87 uniform delivery to multiple animals, are seldom shared (>25%). Finally, details relating to experimental  
88 validation of the delivered radiation dose through some form of dosimeter (ideally traceable to national  
89 standards), are nearly always absent (<15%). Reference to published standards or protocols was as low as  
90 1.3% of published studies in Draeger *et al.* (9).

91 As originally illustrated in the report following the NIST workshop, “investigators cannot report  
92 what they have not considered.”(3). When experiments are not described, a reviewer cannot assess what an  
93 author might or might not have considered, or the soundness of the experimental design at all. In some  
94 instances, metastudies reveal that certain publications contained glaring contradictions in their irradiation  
95 protocols. These include descriptions of irradiators that would be incapable of producing the experiment  
96 described in the methods, non-existent radionuclides, or typographical errors (e.g., kV to MV), which  
97 translate into a thousand-fold difference in energy. These findings suggest that many studies may be  
98 erroneously described, even in cases where there is no clear contradiction in the few details available. For  
99 instance, irradiating multiple mice placed at varying distances from the center of the field will often result in

100 a different dose (and therefore a different biological effect) delivered to each, but this will not be apparent in  
101 a manuscript that simply describes the source used and dose prescribed.

102 In a vacuum, a lack of rigor in describing experimental protocols might not imply anything, but  
103 insufficient communication among authors, or a propensity to copy and combine old protocols without  
104 sufficient rigor can be problematic. When combined with repeated reports outlining difficulties in  
105 harmonizing radiation dosimetry across separate laboratories, (2, 7, 8) a more concerning picture emerges.  
106 As stated in the original NIST report, “Few students or researchers using ionizing radiation in biological  
107 research have training in basic radiation physics. This leads to the difficult situation that when ‘one does not  
108 know what one does not know’ dosimetry design and documentation go unaddressed”. And if it is not  
109 described, a journal or grant reviewer cannot assess whether the experiment is sound.

110 One particular concept not initially addressed in the first NIST workshop concerns the definition of  
111 radiation dose. Indeed – not all dose is created equal. An X-ray tube can be calibrated in terms of exposure,  
112 air kerma, or adsorbed dose to water ( $D_w$ ), depending on which correction factors are used. Reference  
113 dosimetry can assume, or not, the presence of backscatter. Or, an investigator can refer to the radiation dose  
114 being delivered to animals under experimental conditions. Each of these descriptions is scientifically sound,  
115 and yet can lead to differences of up to 50% in reported doses. As described in the next section, the use of  
116 Roentgen (exposure) vs dose for  $^{137}\text{Cs}$  exposures will introduce an error of ~11%. Without further  
117 description, it is impossible to know which quantity an investigator is referring to. To address this, the  
118 American Association of Physicists in Medicine (AAPM)’s Task Group 319 is developing new standards on  
119 how to perform and report experimental dosimetry in kilovoltage cabinet irradiators. However, these new  
120 guidelines have yet to be published, and their implementation will not be instantaneous. In the meantime,  
121 investigators should be careful in how they describe their irradiation protocol in scientific communications.

122  
123 **U.S. Efforts in Dosimetry Harmonization**

124 Inconsistent dosimetry was first outlined by the 2011 NIST panel<sup>3</sup>, and follow-up publications in  
125 2016 by Seed, *et al.*, (8) and Pedersen, *et al.* (7) suggested that only one in five laboratories deliver dose  
126 within 5% of a specified target level, with variations up to 42% observed. From 1971 to 1999, the EULEP-  
127 EURADOS initiative performed 8 similar intercomparisons and consistently found errors of up to ~20%,  
128 and found the majority of institutions could not satisfy a  $\leq 10\%$  homogeneity requirement (10). Radiation  
129 biology research often employs orthovoltage X-ray systems and <sup>137</sup>Cs sources where dosimetry is based on  
130 manufacturer recommendations. For X-rays, results may not be traceable to NIST standard beams because  
131 of the differing maximum beam energy and the filtration used for a certain energy. In addition to traceability  
132 issues, low-energy X-rays introduce variability in dosimetry and radiobiological response in multiple ways.  
133 First, low energy bremsstrahlung radiation is more readily absorbed in tissues than are higher energy  
134 photons from isotope irradiators. Second, low-energy X-rays are more readily absorbed by cortical bone,  
135 inducing differential biological effects in bone marrow and different biological responses (11-13).

136 From 2020 to 2023, The NIAID/RNCP program for harmonization of radiation dose evaluated >30  
137 irradiators (X-ray and cesium), and found that upon initial survey, 50% of machines were outside of  $\pm 5\%$  of  
138 the targeted 4 Gy dose, the benchmark for acceptable irradiation, as is shown in Figure 1. Traceable NIST  
139 beams for X-ray beams were used for cabinet X-ray systems by matching half value layers (HVL) for  
140 radiobiological X-ray beams, although kVp and second HVL are much different. Phantoms for total-body,  
141 and partial-body irradiation were established, and dose was measured using LiF:Mg,Ti thermoluminescent  
142 dosimeters (TLDs) calibrated using the NIST-traceable HVL matched X-ray beams.

143 Variations from the 4 Gy target dose ranged from +/-12% to 25%. After consultations and retesting,  
144 results for nearly all institutions were brought within the required 5% of target dose, and doses were  
145 resolved for the remaining institutions after consideration of physics issues. Also, precision mapping of the  
146 radiation field using multiple TLDs was important and permitted the resolution of dose discrepancies that  
147 aligned with variations in dose. To summarize, there are three problems that are frequently encountered at  
148 institutions that must be addressed. First, the correct physics dose quantity and other characterizations

---

<sup>3</sup> <https://www.nist.gov/programs-projects/medical-archive-workshop-radiobiology-dosimetry-standardization>.

149 should be employed, especially for X-ray machines. An example of this that was found with some  
150 radionuclide sources involving a dose discrepancy of ~11 %. The institutions in question were using the  
151 manufacturer-supplied radiation exposure quantity in Roentgens, rather than the absorbed dose to water; and  
152 the correction needed to achieve absorbed dose to water was 11% - matching the observed error perfectly. A  
153 second problem is not understanding the physics or the measurements given; consultation with a medical  
154 physicist would be of great benefit, but not all sites conducting radiation biology research have access to the  
155 support of a certified medical physicist. Of note, not all certified medical physicists are familiar with  
156 irradiation conditions and calibration protocols of preclinical systems, which can be vastly different than  
157 clinical systems. Finally, an inadequate understanding of dose distribution within the radiation field,  
158 especially in cabinet-style orthovoltage X-ray irradiators, is a significant problem. It is important to  
159 understand the isodose plot of the field, and in many cases, rotation of the subject animals is necessary to  
160 average out the dose. The goal of this NIAID-funded program is the harmonization of dose among  
161 institutions regardless of the source used to establish the radiation.

162

### 163 **French Efforts in Dosimetry Harmonization**

164 In France, the Institute of Radiological Protection and Nuclear Safety (IRSN) has a long tradition of  
165 collaboration between dosimetrists and radiobiology researchers. Technical support has always been  
166 provided to define the best irradiation conditions and reference dose rates, whether for gamma irradiators,  
167 conventional X-ray facilities, or for irradiation at reactors (e.g., the SILENE facility in Valduc, France) with  
168 a concern for the traceability of these data. A dose rate in the most appropriate dosimetric quantity and a  
169 homogeneity level are defined for each irradiation configuration. For example, for cell irradiations, a  
170 configuration is defined for a given container, filling volume, and number of containers. It is not only a  
171 question of defining a reference rate in  $D_w$  or in air kerma according to the reference protocols in use (such  
172 as the AAPM's TG-61 or the IAEA TRS-398), but also taking into account factors that can affect the dose  
173 rate, and defining it at the level of the biological material of interest (14). For small animals, whether for  
174 total-body or localized irradiations, specific phantoms are produced using molds. For programs involving

175 study of bone tissue, electron paramagnetic resonance (EPR) spectroscopy, which enables the dose actually  
176 delivered to the bone to be estimated, has been introduced (15). These skills have been regularly made  
177 available to other national or European research institutes, private companies, and European projects such as  
178 RENEb (16). In Figure 2, the accuracy of the reference dosimetry of beams used to establish calibration  
179 curves in biological dosimetry was evaluated with alanine dosimeters calibrated in Dw. Deviations from  
180 reference dose (10 Gy) range from a few % up to 50%, with about 50% of results within +/- 5%. Dose  
181 quantities used, when known by participants, were either air kerma or Dw. The impact of using different  
182 dose quantities and/or beam quality was noticeable in dicentric assays, but had not been accounted for when  
183 analysing results of biological dosimetry for this inter-laboratory comparison. With the arrival of more  
184 complex new small animal X-ray conformal irradiation platforms to the IRSN (SARRP, Xstrahl), and the  
185 increased use of a linear accelerator (LINAC), the need to have a dedicated physicist on a full-time basis  
186 was felt to be important, due to the greater complexity of use and dosimetry aspects. This recommendation  
187 has also resulted in a dedicated research program at IRSN on the latter aspect, notably to develop new  
188 dosimetry approaches (e.g., for small fields), new protocols for dose rate measurement at biological sites,  
189 verification of manufacturer data, and quality control. A similar need among other users of SARRP type of  
190 X-ray platforms has resulted in hiring additional physicists to ensure the optimal performance of these  
191 machines. This small community quickly organized itself to share experiences, protocols, and methods, and  
192 endeavored to set up a program to verify manufacturer data and provide intercomparison and technical  
193 assistance for machines without local physicist expertise, notably in the framework of a new French national  
194 network of irradiation platforms for radiobiology studies (Réseau des Plateformes National d'Irradiation;  
195 ResPlanDir)<sup>4</sup>.

196

### 197 **Proposed New Instructions for Authors for Dosimetry Reporting**

198 Given the concerns outlined above regarding accuracy and reproducibility of radiation dosimetry in  
199 radiobiological research, and in the interest in ensuring rigor and reproducibility of irradiation experiments,

---

<sup>4</sup> <https://www.france-hadron.fr/en/plateforms/features-modalities.html>


200 the authors of this commentary, in agreement with another recent commentary on this subject by Marcu et  
201 al. (17), propose that journals engaged in publishing radiobiology research introduce new ‘Instructions for  
202 Authors’ addressing minimum standards for reporting radiation dosimetry information. Implementing the  
203 recommendations below would also address issues sometimes encountered in successful study replication.

204 Therefore, the following minimum irradiation parameters are suggested:

- 205 • Name of the irradiator (company, country, model)
- 206 • Type of source (radionuclide (e.g. cesium, cobalt), X-ray)
- 207 • Energy (kVp, mA; for X-ray systems)
- 208 • Filtration (for kV X-ray systems, half value layer if known)
- 209 • Radiation dose and dose rate, source to surface distance (SSD), field size,
- 210 • Sample being irradiated, and
- 211 • Irradiation geometry (nature of animal restraint (e.g., pie or other jig), the material of the sample  
212 holder, shielding (if appropriate), orientation to the source, any animal(s)/sample rotation)
- 213 • Dose quantity (e.g., how radiation dose was measured -  $D_w$ , air kerma, midline tissue, etc.)
- 214 • Daily pre-exposure device calibration method
- 215 • Planning and dose calculation method (required for multiple field conformal dose delivery,  
216 particularly orthovoltage small animal conformal dose delivery systems, optional otherwise)
- 217 • Reporting of any in-run dose measurements (if performed)


218

**Figure 1**



**Figure 1.** A) Performance of U.S. institutions with isotope irradiators within NIAID-supported sites; B) Performance of institutions with orthovoltage X-ray irradiators within NIAID-supported sites. The bold horizontal lines in each panel represent the  $\pm 5\%$  cutoffs for acceptable variance from the 4 Gy target dose. Gray bars indicate paired final retesting results for irradiators with initial test doses that were  $\geq \pm 5\%$  of the target dose.

**Figure 2**



**Figure 2:** Doses measured at 21 European facilities with alanine with a calibration in terms of dose in water and normalized to the dose delivered at the French primary laboratory (modified and used with permission from) (16).

## REFERENCES

1. Hurst GS, Mills WA, Conte FP, Upton AC. Principles and techniques of mixed radiation dosimetry; application to acute lethality studies of mice with the cyclotron. *Radiat Res.* 1956;4(1):49-64.
2. Zoetelief J, Broerse JJ, Davies RW, Octave-Prignot M, Rezvani M, Vergara JC, et al. Protocol for X-ray dosimetry in radiobiology. *Int J Radiat Biol.* 2001;77(7):817-35.
3. Desrosiers M, DeWerd L, Deye J, Lindsay P, Murphy MK, Mitch M, et al. The importance of dosimetry standardization in radiobiology. *J Res Natl Inst Stan.* 2013;118.
4. Stricklin DL, VanHorne-Sealy J, Rios CI, Scott Carnell LA, Taliaferro LP. Neutron radiobiology and dosimetry. *Radiat Res.* 2021;195(5):480-96.
5. Stone HB, Bernhard EJ, Coleman CN, Deye J, Capala J, Mitchell JB, et al. Preclinical data on efficacy of 10 drug-radiation combinations: Evaluations, concerns, and recommendations. *Transl Oncol.* 2016;9(1):46-56.
6. Stern W, Zira S, Boyd C, Iliopoulos I, Peiris P, Fedurin M. Alternative technologies meta study report. Laboratory BN: U.S. Department of Energy, Office of Nuclear Energy. 2022; Report No.: BNL-223583-2022-FORE. <https://doi.org/10.2172/1895064>
7. Pedersen KH, Kunugi KA, Hammer CG, Culberson WS, DeWerd LA. Radiation biology irradiator dose verification survey. *Radiation Research.* 2016;185(2):163-8, 6.
8. Seed TM, Xiao S, Manley N, Nikolich-Zugich J, Pugh J, Van den Brink M, et al. An interlaboratory comparison of dosimetry for a multi-institutional radiobiological research project: Observations, problems, solutions and lessons learned. *Int J Radiat Biol.* 2016;92(2):59-70.
9. Draeger E, Sawant A, Johnstone C, Koger B, Becker S, Vujaskovic Z, et al. A dose of reality: How 20 years of incomplete physics and dosimetry reporting in radiobiology studies may have contributed to the reproducibility crisis. *Int J Radiat Oncol Biol Phys.* 2020;106(2):243-52.
10. Zoetelief J, Broerse JJ, Busscher FA, Hiestand WP, Julius HW, Jansen JT. Recent EULEP dosimetry intercomparisons for whole body irradiation of mice. European Late Effects Project group. *Int J Radiat Biol.* 1997;72(5):627-32.

11. DeWerd LA, Kunugi K. Accurate dosimetry for radiobiology. *Int J Radiat Oncol Biol Phys.* 2021;111(5):e75-e81.
12. Bell BI, Vercellino J, Brodin NP, Velten C, Nanduri LSY, Nagesh PKB, et al. Orthovoltage X-rays exhibit increased efficacy compared with  $\gamma$ -rays in preclinical irradiation. *Cancer Res.* 2022;82(15):2678-91.
13. Poirier Y, Belley MD, Dewhirst MW, Yoshizumic TT, Down JD. Transitioning from gamma rays to X-rays for comparable biomedical research irradiations: Energy matters. *Radiat Res.* 2020;193(6):506-11.
14. Dos Santos M, Paget V, Trompier F, Gruel G, Milliat F. Dosimetry for cell irradiation using orthovoltage (40-300 kV) X-ray facilities. *J Vis Exp.* 2021;10.3791/61645(168).
15. Guillou M, L'Homme B, Trompier F, Gruel G, Prezado Y, Dos Santos M. Preclinical modeling of low energy X-rays radiological burn: Dosimetry study by monte carlo simulations and EPR spectroscopy. *Front Physiol.* 2022;13:1075665.
16. Trompier F, Baumann M, Barrios L, Gregoire E, Abend M, Ainsbury E, et al. Investigation of the influence of calibration practices on cytogenetic laboratory performance for dose estimation. *Int J Radiat Biol.* 2017;93(1):118-26.
17. Marcu LG, Bezak E, Sykes PJ. Radiation research journals need to stipulate minimal dosimetry requirements for publishing research using X-radiation exposures. *Radiat Res.* 2022;198(2):204-6.